

**Test Report No. 7191303172-EEC23-WBH**  
dated 24 Feb 2023



PSB Singapore

**Add value.  
Inspire trust.**

**Note:** This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

**SUBJECT**

Testing of Gloves

**CLIENT**

Zhonghong Pulin Medical Products Co., Ltd.  
West Industrial Park, BDA, Luannan,  
Tangshan, 063500 Hebei,  
P.R. China

**SAMPLE SUBMISSION DATE**

09 Feb 2023

**TEST DATE**

10 Feb 2023 to 15 Feb 2023

**DESCRIPTION OF SAMPLES**

S/N	Product Description	Brand	Colour	Lot No.	Size	Sample Received (pieces)	Manufactured by
1	Nitrile Exam Gloves	Nil	Blue	Nil	XS	220	Zhonghong Pulin Medical Products Co., Ltd.
2					S	220	
3					M	220	
4					L	220	
5					XL	220	
6					XXL	220	

Lot size as specified by client: 35,001 to 150,000 pieces per lot

**METHOD OF TEST**

EN 455-1:2020+A1:2022 Medical gloves for single use  
Part 1: Requirements and testing for freedom from holes



		LA-2007-0380-A	LA-2007-0386-C	The results reported herein have been performed in accordance with the terms of accreditation under the Singapore Accreditation Council. Inspections/Calibrations/Tests marked "Not SAC-SINGLAS Accredited" in this Report are not included in the SAC-SINGLAS Accreditation Schedule for our inspection body/laboratory.
		LA-2007-0381-F	LA-2010-0464-D	
LA-2007-0382-B	LA-2018-0702-B			
LA-2007-0383-G	LA-2018-0703-G			
LA-2007-0384-G	LA-2020-0747-L			
LA-2007-0385-E				

Laboratory:  
TUV SÜD PSB Pte. Ltd.  
15 International Business Park  
TUV SÜD @ IBP  
Singapore 609937

Phone : +65-6778 7777  
E-mail: info.sg@tuvsud.com  
<https://www.tuvsud.com/sg>  
Co. Reg : 199002667R

Regional Head Office:  
TUV SÜD Asia Pacific Pte. Ltd.  
15 International Business Park  
TUV SÜD @ IBP  
Singapore 609937  
**TUV®**

**Test Report No. 7191303172-EEC23-WBH**  
dated 24 Feb 2023



PSB Singapore

**RESULTS**

Sample: Nitrile Exam Gloves, Blue

Table: Results for EN 455-1:2020+A1:2022

Single sampling plan, general inspection level I, AQL 1.5

Clause	Tests	Size	Requirements	No. of non-conformances allowed (pieces)	Number tested (pieces)	Actual no. of non-conformances found (pieces)	Inferred results
4 5	Freedom from holes	XS	Shall not leak	7	200	2	Passed
		S		7	200	4	Passed
		M		7	200	0	Passed
		L		7	200	1	Passed
		XL		7	200	1	Passed
		XXL		7	200	4	Passed

**REMARK**

Brand and Lot No. were not provided by client.

  
\_\_\_\_\_  
Yeo Poh Kwang  
Test Officer



  
\_\_\_\_\_  
Wong Bee Hui  
Product Manager  
Non-Asset Intensive Laboratories  
Product Service

**PHOTO**



Photo: Nitrile Exam Gloves, Blue



**Test Report No. 7191303172-EEC23-WBH**  
**dated 24 Feb 2023**



Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, 15 International Business Park TÜV SÜD @ IBP Singapore 609937.
6. The tests carried out by TÜV SÜD PSB and this report are subject to TÜV SÜD PSB's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SÜD Group.

Effective 26 January 2021

